
Subject: Re: Plasma equivalents NFHS-4 and NFHS-5
Posted by fred.arnold@icf.com on Wed, 06 Jul 2022 17:52:10 GMT
[View Forum Message](#) <> [Reply to Message](#)

In NFHS-5 and NFHS-4, the blood glucose test results are expressed in mg/dl. In both surveys, capillary blood was collected from a finger stick to measure non-fasting whole blood glucose using the FreeStyle Optium H glucometer with glucose test strips. The survey reports do not have plasma equivalents for the test results.
